



Consultant
Training Centre
Calibration Services

CERTIFICATE OF CALIBRATION

Page 1 of 2

Instrument Details:

Name : Gas Detector
Manufacturer : Honeywell BW
Type/Model : MCXL-XWHM-Y-CN
Serial Number/Code : KA424-8083463
Capacity/Nominal/Range : Oksigen (O₂) 18 %
Methane (CH₄) 50 % LEL
Carbon Monoksida (CO) 100 ppm
Hydrogen Sulfide (H₂S) 25 ppm
Resolution : 0,1 %
1 % LEL
1 ppm
0,1 ppm

Standard & Equipment:

Code : BMD-KMA-GAS01
Traceable to SI through : CALGAZ
Reference Method : Honeywell Calibration Handbook
Calibration Procedure : BMD/IK/7.2/KM/11

Testing Condition:

Location : In Lab
Date of Calibration : June 24, 2026
Temperature (°C) : 26,2 ± 1,2
Humidity (%) : 40 ± 3,6

Owner's Identification:


Name : PT. ELNUSA, Tbk
Address : Warehouse Duri.
Jl. Rangau KM 3.5 Pematang Pudu, Kec. Mandau, Bengkalis – Riau
Order Number : BMD-06390626
Unit Received : June 8, 2026

Note:

Calibration report and attachment available on next pages is an integral part of this certificate

Date of issue:

Tangerang Selatan, June 30, 2026



Risky Elsa Putri Amelia, S.Pd

Director

Certificate Number:

017.0626/AU12